

# Performance Evaluation of a New Culture Colorimetric Detection Assay

Cartesio Favalli<sup>1,2,3</sup> , Marco Favaro<sup>1</sup> , Flavia Santi<sup>4</sup> , Micol Piperno<sup>1</sup> , Cartesio D'Agostini<sup>1,3</sup> , Marco Ciotti<sup>1</sup> 



**Cite this article as:** Favalli C, Favaro M, Santi F, Piperno M, D'Agostini C, Ciotti M. Performance Evaluation of a New Culture Colorimetric Detection Assay. *Eurasian J Med* 2019; 51(1): 5-7.

**ORCID IDs of the authors:**  
 C.F: 0000-0002-6246-2628  
 M.F: 0000-0002-3502-8332  
 F.S: 0000-0001-6150-0381  
 M.P: 0000-0002-2120-370X  
 C.D: 0000-0001-9199-3340  
 M.C.: 0000-0002-9943-9130

<sup>1</sup>Department of Clinical Microbiology and Virology, Polyclinic Tor Vergata Foundation, Rome, Italy

<sup>2</sup>Department of Biochemical Sciences, Catholic University "Our Lady of Good Counsel", Tirana, Albania

<sup>3</sup>Department of Experimental Medicine and Surgery, University of Rome "Tor Vergata", Rome, Italy

<sup>4</sup>Department of Statistical Sciences, "La Sapienza" University of Rome, Rome, Italy

Received: November 9, 2017

Accepted: June 27, 2018

Available Online Date: November 30, 2018

Correspondence to: Marco Ciotti  
 E-mail: marco.ciotti@ptvonline.it

DOI 10.5152/eurasianjmed.2018.17364

©Copyright 2019 by the Atatürk University School of Medicine - Available online at [www.eurasianjmed.com](http://www.eurasianjmed.com)

## ABSTRACT

**Objective:** To evaluate the performance of the culture colorimetric detection assay MYCO WELL D-ONE® (MWD-ONE), designed to detect sexually transmitted infections using real-time polymerase chain reaction (PCR) as a reference method.

**Materials and Methods:** One hundred and ten urogenital samples were screened for *Gardnerella vaginalis* (GV), *Trichomonas vaginalis* (TV), *Mycoplasma hominis* (MH), *Mycoplasma spp.*, and *Ureaplasma urealyticum* (UU)/*Ureaplasma parvum* (UP) using the MWD-ONE and real-time PCR assays *Gardnerella vaginalis*/Lactobacillus species Real-TM Quant and Anyplex II STI-7 Detection, respectively.

**Results:** GV was detected in 33 samples by both the MWD-ONE and real-time PCR, while 6 samples gave discordant results. TV was detected by both MWD-ONE and Anyplex II STI-7 Detection kits in 3 samples, while 107 were negative. MH was detected by both methods in 5 cases, 4 samples gave discordant results, and 101 were negative. *Mycoplasma genitalium* (MG) was detected by Anyplex II STI-7 in 2 cases, 1 of which was detected as *Mycoplasma spp.* by MWD-ONE. Ten samples were positive by MWD-ONE, and 98 were negative with both assays. With regard to UU/UP, 24 cases were detected by MWD-ONE and Anyplex PCR, 25 by PCR only, 4 by MWD-ONE, and 57 tested negative with both methods.

The positive predictive values (PPV) and negative predictive values (NPV) of the MWD-ONE assay for the pathogens tested were as following: GV, PPV 94.3%, NPV 94.7%; TV, PPV and NPV 100%; MH, PPV 71.4%, NPV 98.1%; *Mycoplasma spp.*, PPV 9.1%, NPV 98.9%; and *Ureaplasma spp.*, PPV 85.7 %, NPV 69.5 %. The agreement between the MWD-ONE and PCR was strong for GV and MH (k=0.8 and 0.7, respectively); perfect for TV (k=1); and moderate for UU/UP (k=0.4).

**Conclusion:** MWD-ONE assay appears to be suitable for routine testing of sexually transmitted infections.

**Keywords:** MYCO WELL D-ONE, sexually transmitted infections, colorimetric detection assay, real-time PCR

## Introduction

Sexually transmitted infections (STI) are among the most common acute conditions in the human population worldwide. They have a great impact on sexual and human health and represent an important health burden [1]. There are over 30 infections that can be sexually transmitted, caused by bacteria, viruses, and parasites. Most of these infections are asymptomatic or have mild symptoms, and if unrecognized and left untreated, they may have serious consequences. In this work, we focused on some of sexually transmitted pathogens: *Gardnerella vaginalis* (GV), *Trichomonas vaginalis* (TV), *Mycoplasma hominis* (MH), *Mycoplasma genitalium* (MG), *Ureaplasma urealyticum* (UU), and *Ureaplasma parvum* (UP). These pathogens have been associated with clinical conditions such as bacterial vaginosis, pelvic inflammatory disease, urethritis, cervicitis, infertility, ectopic pregnancy, and adverse pregnancy outcomes [2–7].

Currently, real-time polymerase chain reaction (PCR) is the method of choice to detect these microorganisms because of its high sensitivity, specificity, and capability to search for multiple targets [8–11]. However, in low-resource settings, molecular methods are not always available, and less expensive methods may be required. Colorimetric culture assays based on the detection of live pathogens through enzymatic reactions may be a valid alternative [12]. In this study, we tested the performance of a new colorimetric detection assay, MYCO WELL D-ONE, which detects GV, TV, MH, MG, UU, and UP, and the results were compared with those obtained by using Anyplex II STI-7 Detection kit for TV, MH, MG, UU, and UP, and *Gardnerella vaginalis*/Lactobacillus species Real-TM Quant assay for GV.

## Materials and Methods

### Study Group

One hundred and ten outpatients with a suspected urogenital infection were admitted to the laboratory of Clinical Microbiology, Polyclinic Tor Vergata, Rome, Italy, and screened for sexually transmitted pathogens including GV, CT, NG, TV, MH, MG, UU, and UP. The number of obtained biological samples was as follows: 70 cervical, 30 vaginal, and 8 urethral swabs, and 2 seminal liquid, which were stored at +4°C until analysis. The study was carried on residual samples, otherwise destroyed, following a routine diagnostic analysis. The identification number of the patient was converted into an alphanumeric code making anonymous all information regarding the patient. Therefore, no informed consent was required. The Polyclinic Tor Vergata Foundation Ethics Committee approved the study (Protocol R.S. 152/17).

### Detection of GV, TV, MH, MG, UU, and UP by the MYCO WELL D-ONE Assay

The urogenital samples were processed with the MYCO WELL D-ONE assay following the manufacturer's instructions (CPM SAS, Formello, Italy). Samples were suspended in sterile saline solution for 3–5 minutes until obtaining a homogeneous suspension. Then, 150 µl of the obtained suspension were added to each well plate (wells 1 to 32). All wells were covered with two drops of sterile paraffin, except wells 27 and 31, which were dedicated to the microscopic observation (40X). A series of antibiotics (levofloxacin, tetracycline, moxifloxacin, erythromycin, and clindamycin) at different concentrations were inoculated in the wells 9 through 24 according to the Clinical & Laboratory Standards Institute guidelines [13]. Then, the samples were incubated at 36±1°C for 24–48 h.

### Detection of GV by the Gardnerella vaginalis/Lactobacillus Species Real-TM Quant Assay

*Gardnerella vaginalis* was detected by real-time PCR using the Gardnerella vaginalis/Lactobacillus species Real-TM Quant assay (Sacace Biotechnologies, Como, Italy), which targets the 16S rRNA region of both GV and Lactobacillus species. The PCR amplification was performed according to the manufacturer's instructions.

### Detection of TV, MH, MG, UU, and UP by Anyplex II STI-7 Detection

Anyplex II STI-7 Detection (Seegene, Seoul, Korea) is a multiplex real-time PCR assay that enables detection of the seven major causative agents of sexually transmitted infections (CT, NG, TV, MH, MG, UU, and UP) in a single reaction tube. The assay is based on the Dual-Priming Oligonucleotides and Tagging Oligonu-

cleotide Cleavage and Extension technology, as already described [11].

### Statistical Analysis

The performance of MYCO WELL D-ONE test was compared with that of the real-time PCR, which was used as the gold standard. Sensitivity, specificity, positive predictive values (PPV) and negative predictive values (NPV) were calculated for GV, TV, MH, MG, UU, and UP. Confidence intervals at the 95% were estimated assuming exact binomial distribution. The agreement between the two tests was evaluated by the Cohen's kappa statistics. All analyses were performed using the STATA software version 11 (Stata Corporation, College Station, TX).

### Results

Of the 110 biological samples tested, GV was detected by both methods in 33 samples (30%);

2 samples (1.81 %) were positive with MWD-ONE, 4 samples (3.6%) with GV Real-TM Quant real-time PCR, and 71 samples were negative with both methods.

*Trichomonas vaginalis* was detected by MWD-ONE and Anyplex PCR in 3 out of 110 (2.73%) samples. The remaining samples were all negative.

*Mycoplasma hominis* (MH) was detected in 2 out of 110 (1.81%) only with Anyplex PCR and in 2 out of 110 (1.81%) only with MWD-ONE. A concordant result was obtained in five cases. One hundred and one samples tested negative with both methods.

*Mycoplasma genitalium* (MG) was detected by Anyplex II STI-7 in 2/110 (1.81%), one of which

was also detected by MWD-ONE as *Mycoplasma* species because the latter assay cannot discriminate between MG and *Mycoplasma* species. Ten more samples positive with MWD-ONE tested negative with Anyplex II STI-7 assay, suggesting the presence of mycoplasmas different from MG. The remaining 98 samples gave a negative result with both methods.

The Anyplex assay detected UU in 6 out of 110 (5.45%), UP in 29 out of 110 (26.3%) and UU+UP in 14 out of 110 (12.7%), respectively. Four samples were positive with MWD only, and 57 were negative with both assays. Results are summarized in Table 1.

Sensitivity, specificity, PPV, and NPV of MYCO WELL D-ONE assay toward GV, TV, MH, *Msp*, and *Usp*, were calculated using the real-time PCR as the gold standard. Results are reported in Table 2.

The level of agreement between the two tests measured by Cohen's kappa statistics was strong for GV (k=0.8) and MH (k=0.7), perfect for TV (k=1), and moderate for UU/UP (k=0.4) (Table 2).

### Discussion

Laboratory diagnosis of STI is crucial for the choice of appropriate therapy and for preventing the sequelae induced by the transmitted pathogen. Over the years, several diagnostic tests have been developed and used to detect STI: culture methods, serological and hybridization assays, and antigen detection [14–18]. In this study, we evaluated the performance of a new culture colorimetric detection assay named MYCO WELL D-ONE, that detects some of the sexually transmitted pathogens, namely: GV,

**Table 1. Pathogens Detection by PCR and MYCO WELL D-ONE**

	*M+/PCR+	M+/PCR-	M-/PCR+	M-/PCR-	Total	Cohen's kappa
<i>Gardnerella vaginalis</i>	33	2	4	71	110	0.8
<i>Trichomonas vaginalis</i>	3	-	-	107	110	1
<i>Mycoplasma hominis</i>	5	2	2	101	110	0.7
<i>Mycoplasma genitalium</i>	1	-	1	98	-	
† <i>Mycoplasma spp.</i>	-	10	-	-	110	
<i>Ureaplasma urealyticum</i>	3	-	3	-	-	
<i>Ureaplasma parvum</i>	8	-	21	-	-	
‡ <i>Ureaplasma urealyticum</i> + <i>Ureaplasma parvum</i>	13	4	1	57		
<b>Total</b>	<b>24</b>	<b>4</b>	<b>25</b>	<b>57</b>	<b>110</b>	<b>0.4</b>

\*M: MYCO WELL D-ONE  
 †Msp: Sector *Mycoplasma spp.* of MYCO WELL D-ONE (including eventually growth of *M.genitalium*, *M.fermentans*, *M.pneumoniae*, *M.pirum*, and other *M.spp.*, except *M.hominis*).  
 ‡Usp: Sector of *Ureaplasma spp.* of MYCO WELL D-ONE (including *U.urealyticum* and *U.parvum*).

**Table 2.** Sensitivity, Specificity, Positive Predictive Values (PPV), and Negative Predictive Values (NPV) of the MYCO WELL D-ONE Assay

Sexually Transmitted Pathogens	Prevalence	Sensitivity	Specificity	Positive Predictive Value (NPV)	Negative Predictive Value (NPV)
<i>Gardnerella vaginalis</i> (N=110)	33.6%	89.2% (74.6–97.0)	97.3% (90.4–99.7)	94.3% (80.7–98.5)	94.7% (87.5–97.8)
<i>Trichomonas vaginalis</i> (N=110)	2.7%	100% (29.2–100)	100% (96.6–100)	100%	100%
<i>Mycoplasma hominis</i> (N=110)	6.4%	71.4% (29–96.3)	98.1% (92.4–99.7)	71.4% (29.0–96.3)	98.1% (92.4–99.7)
<i>Mycoplasma spp.</i> (N=110)	1.8%	50.0% (1.3–98.8)	90.7% (83.6–95.5)	9.1% (2.17–31.1)	99.0% (96.1–99.7)
<i>Ureaplasma urealyticum</i> + <i>Ureaplasma parvum</i> (N=110)	44.55%	48.98% (34.4–63.7)	93.4% (84.0–98.18)	85.71% (65–94.1)	69.5% (62.2–75.1)

TV, MH, MG, UU, and UP. The last three pathogens (MG, UU, and UP) are detected, but not discriminated by this assay. Therefore, a colorimetric change toward the positivity in the well containing selective medium for the growth of mycoplasmas can be attributed not only to MG, but also to other *Mycoplasma spp.* such as *Mycoplasma fermentans*, *Mycoplasma penetrans*, or others. This would explain the discrepant results observed between the colorimetric assay and the real-time PCR. The same could apply to some discordant results observed in the case of GV (2) and UU/UP (4) detection. It can be envisaged that microorganisms present in the sample analyzed and characterized by similar biochemical requirements may cause a pH change in the well that is misinterpreted as positivity. This is a limitation of the assay because it requires further molecular testing to identify the pathogen involved.

Although PCR is currently the method of choice in the most-advanced microbiology laboratories, it is not always available in low-resource settings. In this contest, colorimetric culture methods can be a valid alternative [12]. Actually, despite the limitations reported above, the overall performance of the test was quite good. Cohen's kappa statistics revealed a strong agreement between the MWD-ONE and PCR for GV and MH, a perfect agreement for TV, and a moderate agreement for UU/UP. Furthermore, MWD-ONE showed a high NPV for all the microorganisms investigated, except UU/UP, making it suitable for routine testing of sexually transmitted pathogens.

In conclusion, although PCR is more sensitive and specific, MWD-ONE is an economical and practical system that could be appealing for laboratories equipped for general microbiology

purposes, especially in low-resource settings. In addition, the assay offers the possibility to perform antimicrobial susceptibility testing in the same plate.

**Ethics Committee Approval:** Ethics committee approval was received for this study from the Ethics Committee of Polyclinic Tor Vergata Foundation.

**Informed Consent:** N/A

**Peer-review:** Externally peer-reviewed.

**Author Contributions:** Concept – F.C.; Design – F.C., F.M.; Supervision – F.C., C.M.; Resources – CPM SAS; Materials – Laboratory of Clinical Microbiology, CPM SAS; Consultants; Data Collection and/or Processing – F.M., P.M., D.C.; Analysis and/or Interpretation – S.F., F.C., C.M.; Literature Search – C.M., F.C.; Writing Manuscript – C.M.; Critical Review – F.C.

**Acknowledgements:** The authors would like to thank CPM SAS for providing the reagents.

**Conflict of Interest:** The authors declared no conflicts of interest.

**Financial Disclosure:** The authors declare that this study has received no financial support.

## References

- Newman L, Rowley J, Hoorn SV, et al. Global estimates of the prevalence and incidence of four curable sexually transmitted infections in 2012 based on systematic review and global reporting. *PLoS One* 2015; 10: e0143304.
- Machado A, Cerca N. Influence of biofilm formation by *Gardnerella vaginalis* and other anaerobes on bacterial vaginosis. *J Infect Dis* 2015; 15: 1856-61. [CrossRef]
- Moodley P, Wilkinson D, Connolly C, Moodley J, Sturm AW. *Trichomonas vaginalis* is associated with pelvic inflammatory disease in women infected with human immunodeficiency virus. *Clin Infect Dis* 2002; 34: 519-22. [CrossRef]

- Deguchi T, Yoshida T, Miyazawa T, et al. Association of *Ureaplasma urealyticum* (biovar 2) with nongonococcal urethritis. *Sex Transm Dis* 2004; 31: 192-5. [CrossRef]
- Volgmann T, Ohlinger R, Panzig B. *Ureaplasma urealyticum*—harmless commensal or underestimated enemy of human reproduction? A review. *Arch Gynecol Obstet* 2005; 273: 133-9. [CrossRef]
- Waites KB, Katz B, Schelonka RL. Mycoplasmas and ureaplasmas as neonatal pathogens. *Clin Microbiol Rev* 2005; 18: 757-89. [CrossRef]
- Larsen B, Hwang J. Mycoplasma, Ureaplasma, and adverse pregnancy outcomes: a fresh look. *Infect Dis Obstet Gynecol* 2010; 2010: pii: 521921.
- Lee SJ, Park DC, Lee DS, Choe HS, Cho YH. Evaluation of Seeplex(R) STD6 ACE Detection kit for the diagnosis of six bacterial sexually transmitted infections. *J Infect Chemother* 2012; 18: 494-500. [CrossRef]
- Choe HS, Lee DS, Lee SJ, et al. Performance of AnyplexTM II multiplex real-time PCR for the diagnosis of seven sexually transmitted infections: comparison with currently available methods. *Int J Infect Dis* 2013; 17: e1134-40.
- Kim Y, Kim J, Lee KA. Prevalence of sexually transmitted infections among healthy Korean women: Implications of multiplex PCR pathogen detection on antibiotic therapy. *J Infect Chemother* 2014; 20: 74-6. [CrossRef]
- Camporiondo MP, Farchi F, Ciccozzi M, et al. Detection of HPV and co-infecting pathogens in healthy Italian women by multiplex real-time PCR. *Infez Med* 2016; 24: 12-7.
- Boum Y 2nd, Oriquiriza P, Rojas-Ponce G, et al. Use of colorimetric culture methods for detection of *Mycobacterium tuberculosis* complex isolates from sputum samples in resource-limited settings. *J Clin Microbiol* 2013; 51: 2273-79. [CrossRef]
- Clinical and Laboratory Standards Institute (CLSI). Performance Standards for Antimicrobial Susceptibility Testing. 27th ed. CLSI supplement M100. Wayne, PA: Clinical and Laboratory Standards Institute; 2017.
- Jensen JS, Hansen HT, Lind K. Isolation of *Mycoplasma genitalium* strains from the male urethra. *J Clin Microbiol* 1996; 34: 286-91.
- Baseman JB, Cagle M, Korte JE, et al. Diagnostic assessment of *Mycoplasma genitalium* in culture-positive women. *J Clin Microbiol* 2004; 42: 203-11. [CrossRef]
- Hamasuna R, Osada Y, Jensen JS. Isolation of *Mycoplasma genitalium* from first-void urine specimens by coculture with Vero cells. *J Clin Microbiol* 2007; 45: 847-50. [CrossRef]
- Edberg A, Aronsson F, Johansson E, Wikander E, Ahlqvist T, Fredlund H. Endocervical swabs transported in first void urine as combined specimens in the detection of *Mycoplasma genitalium* by real-time PCR. *J Med Microbiol* 2009; 58: 117-20. [CrossRef]
- Abdelrahim NA, Ahmed HI, Fadl-Elmula IM, Bayoumi MA, Homeida MM. Sexually transmitted infections other than HIV/AIDS among women of low socio-economic class attending antenatal clinics in Khartoum, Sudan. *Int J STD AIDS* 2017; 28: 781-87. [CrossRef]